

		EQUIPMENT VIOLATION LIST – ARTICLE 175	
Original	Revised		
2	1	301	Operator did not use all practicable procedures to achieve occupational doses that are as low as reasonably achievable (ALARA) below the limits specified in this code 175.03 (b) (l) (i)
2	1	302	Operator did not use all practicable procedures to achieve public doses that are as low as reasonably achievable (ALARA) below the limits specified in this code 175.03 (b) (l) (i)
2	1	303	Operator did not implement a radiation protection program sufficient to ensure compliance with the provisions of this code 175.03 (b) (l) (ii)
2	1	304	Operator did not provide a radiation safety officer to ensure the implementation of a radiation protection program 175.03 (b) (l) (iii)
3	1	305	Operator did not provide for a radiation safety committee to administer the radiation protection program and to oversee the administration of a quality assurance program 175.03 (b) (l) (iv)
1	1	306	Operator did not provide a quality assurance program for the use of radiation – producing equipment 175.03 (b) (l) (v)
2	1	307	Operator did not ensure that all personnel involved in the use of radiation –producing equipment are supervised, are instructed and are competent to safely use such radiation sources 175.03 (b) (l) (vi)
2	1	308	Operator did not ensure that acceptance testing was performed on all diagnostic equipment before the first use of such equipment on humans 175.03 (b) (l) (viii)
1	1	309	Registrant did not control the occupational dose to an individual to ensure that such dose does not exceed an annual effective dose equivalent of 0.5sv (5rem) 175.03 (c) (l) (i)
1	1	310	Registrant did not conduct operations so that the dose in any unrestricted area from external doses does not exceed 0.02mSv (0.002rem) in any one hour 175.03 (d) (l) (i)
2	2	311	Registrant did not cause to be made all surveys that are necessary to comply with this code and / or are necessary to evaluate radiation levels and potential radiological hazards 175.03 (f) (l)
2	1	312	Personnel monitoring devices not supplied for individuals here required by this code 175.03 (f) (2)

2	1	313	Registrant removed an exposure (s) from an individuals exposure record without prior authorization from the Department of Health 175.03 (f) (2) (iv)
3	3	314	Records of the radiation protection program including audits and / or other reviews of program content and implementation not retained by registrant for three (3) years after the record is made 175.03 (k) (2)
3	2	315	Records of surveys and / or calibrations required by this code not retained by registrant for a period of three (3) years after the record is made 175.03 (k) (4) (i)
2	2	316	Registrant did not retain the records of prior occupational dose and exposure history until the Department authorized their disposition 175.03 (k) (6)
2	2	317	Registrant did not maintain records of doses received by all individuals for
1	1	318	Registrant did not immediately report to the Department each event involving a source of radiation that may have caused an individual to receive a total effective dose equivalent of 0.25 Sv (25rem) or more 175.03l (2) (i)
2	1	319	Registrant did not report to the Department within 24 hours of discovery, each event that may have cause an individual to receive, in a period of 24 hours, a total effective dose equivalent exceeding 0.05Sv (5rem) 175.03l (2) (ii)
2	2	320	Registrant did not submit a written report to the Department within thirty (30) days after any occurrence listed in this section of the code 175.03l (3)
2	1	321	Registrant did not notify individual involved of an occurrence which required a written report to be submitted to the Department under 175.03l (3) – 175.03l (6)
1	1	322	Individual did not adhere to rules, regulations or orders imposed by the Department to protect the public health and safety from radiation hazards 175.03 (n) (1)

2			
3	2	401	Copy of Article 175 of the New York City Health Code is not posted or readily available 175.04 (b) (1) and (2)
3	2	402	A current copy of the “Notice to Employees” Prescribed by the Department is not posted 175.04 (b) (3)

2	2	403	Radiation worker not advised annually of the worker's exposure to radiation as shown in records maintained by the registrant pursuant to 175.03 (k) (8) and 175.04 (c) (2)
1	1	404	Facility did not afford the Department at all reasonable times an opportunity to inspect the radiation source (s) and the installation at which such source (s) is (are) located and each record required to be maintained by this Article 175.04 (e)
1	1	405	Facility did interfere with or obstruct Department personnel in carrying out an inspection, survey or examination in performance of duties for the Department of Health 3.15
1	1	406	Department inspectors were unable to consult privately with workers concerning matters of occupational radiation protection or other matters related to applicable provisions of this code 175.04 (f) (1)

PROFESSIONAL PRACTITIONERS AND RELATED PROVISIONS 175.06

1	1	601	Person other than a professional practitioner did direct or order the application of radiation to a human being 175.06 (c)
1/2	1	602	Professional practitioner who directed or ordered the application of radiation to a human being did not comply with the provisions of the license of the practitioner under the Education Law of the State of New York and / or applicable provisions of part 89 of Title 10 of the Codes, Rules and regulations of the State of New York and / or Article 35 of the Public Health Law of State of New York 175.06 (c) (1) and (2)
2	1	603	Radiologic technologist did not comply with the requirements of the technologist's license 175.06 (d)

QUALITY ASSURANCE PROGRAMS 175.07

1	1	701	No quality assurance program implemented 175.07 (b)
1	1	702	No quality assurance sensitometry tools available at the facility 175.07 (b)
1		703	Sensitometry not performed or grossly inadequate 175.07 (b)
1	1	704	No processor quality assurance program in the effect (a lapse of five (5) or more consecutive days in determining daily sensitometry data when radiography has been conducted in the last three (3) months 175.07 (b)
1	1	705	Sensitometry values exceeded tolerance limits for a period of one week or longer, without corrective action being taken by the facility 175.07 (b)
2	1	706	Quality Assurance Manual not provided or incomplete 175.07 (b) (I) (i)

2	1	707	Radiographic and fluoroscopic equipment test not performed at the prescribed times 175.07 (b) (l) (ii)
2	1	708	Equipment records not maintained 175.07 (b) (l) (iii)
2	1	0708A	Equipment records are incomplete 175.09 (b) (i) (iii)
2	1	709	Quality assurance training program for employees is inadequate or nonexistent 175.07 (b) (i) (iv)
2/1	1	710	Radiation safety training program for employees is inadequate or nonexistent 175.07 (b) (i) (iv)
2	1	711	Output measurements (entrance skin exposure) not made at the prescribed cycle 175.07 (b) (l) (v)
3	2	712	Repeat / Reject analysis not performed 175.07 (b) (i) (ix)

REGISTRATION AND INSPECTION; PERMITS 175.51

1	1	5101	Owner or operator did not obtain a current certificate of registration prior to establishing, maintaining or operating a radiation installation 175.51 (b) (l)
1	1	5102	Owner or operator did not obtain a current certificate of registration prior to installing any radiation equipment in operable condition 175.51 (b) (l)
1	1	5103	New facility did not apply for registration at least thirty (30) days before establishing a radiation installation and / or installing x-ray equipment 175.51 (d) (l)
1	1	5104	Facility did not apply for new registration at least thirty (30) days prior to a change in operator or location 175.51 (d) (2)
1	1	5105	Facility did not apply for renewal of registration at least thirty (30) days prior to the expiration of the certificate of registration. 175.51 (e)
2	1	5106	Current certificate of registration not posted 175.51 (k)
1	1	5107	Operator of radiation installation did not report to the Department within ten (10) days a change affecting the registration 175.51 (M)
1	1	5108	Facility did not afford the Department, at a reasonable time, an opportunity to inspect the radiation installation at a frequency specified in 175.51n (2) and (3). 175.51n (l)
2	1	5109	Radiation installation did not maintain radiation exposures as far below the limits set forth in this article as practicable 175.51 (n) (l)
1	1	5110	Person did not engage in the business of selling new or used radiation equipment and / or assembling, installing or repairing such equipment without a permit therefore issued by the Department 175.51 (o) (l)

1	1	5111	Person engaged in the business of selling new or used radiation equipment and / or assembling, installing or repairing such equipment did not comply with the reporting requirements in the Health code
----------	----------	-------------	---

PROHIBITED USED AND ACTIVITIES 175.53
--

1	1	5301	Non- image intensified fluoroscope used 175.53 (a) (6)
1	1	5302	Individual not authorized to operate x-ray equipment pursuant to New York State Law did operate such equipment 175.53 (b) (l)
1	1	5303	X-ray or fluoroscopic equipment sold, leased transferred or loaned to persons not engaged in an occupation where such use is permitted 175.53 (b) (2)
1/2	1	5304	Radiation – producing equipment sold, leased, transferred or loaned which when place into operation does not meet the requirements of Article 175. 175.53 (b) (3)
1/2	1	5305	Radiation – producing equipment repaired such that when placed into operation it does not meet the requirements of Article 175.53 (b) (3)

SURVEYS, SHIELDING REQUIREMENTS AND OPERATOR PROTECTION 175.54

2	1	5401	Radiation protection survey by a qualified expert has not been performed 175.54 (b)
3	1	5402	Radiation protection survey, reportedly performed, by a qualified expert has not been submitted to the Department for review 175.54 (b)
3	1	5403	Radiation protection survey, reportedly performed, by a qualified expert, not maintained by the owner or operator as part of the installation’s radiation records 175.54 (b) (fixed radiographic, installations except dental, mobile, portable or podiatric)
2	2	5404	Exposure switch is operable outside the shielded area (control console not located within a structure so constructed that radiation has to be scattered at least twice before entering the structure or where structure is provided with an interlocked door) 175.54 (c) (l) A (a) and (b)
2	1	5405	Shield used for operator protection is not 2.13m (7feet) in height 175.54 (c) (l) (i) (B)
2	1	5406	Operator not provided with an adequate viewing system to observe the patient during any exposure at the operator’s protected position 175.54 (c) (l) (i) (B)

2	2	5407	Operator not provided with a viewing system to observe any entry into the radiographic room at the operator's protected position 175.54 (c) (I) (i) (B)
2	1	5408	Operator not provided with an adequate means of communication with the patient from the operator's position 175.54 (c) (I) (i) (C)
2	1	5409	Exposure control located in a control console lacking an interlocked door is not placed at least 1.02m (40in) from all open edges of the structures where radiation may enter 175.54 (c) (I) (i) (E)
2	1	ADD	The operator did allot unobstructive for space at the control console 175.54 (c) (i) (d)

MOBILE, PORTABLE,PODIATRIC, DENTAL INSTALLATIONS (EXCLUDING MAMMOGRAPHY)

1	1	5410	Operator not provided with the means to initiate an x-ray exposure at least 2m (6 feet) from the patient or behind a protective barrier 175.54 (c) (2) (i)
2	1	5411	Mobile and portable x-ray systems used as fixed radiographic installations do not meet the operator protection standards of 175.54 (c) (I) (i) (A) or (B)
2	1	5412	Operator of a mobile or portable x-ray unit not provided with personnel monitoring 175.54 (c) (2) (ii)
2	1	5413	Operator of a mobile or portable x-ray unit not provided with a protective apron of at least 0.25mm lead equivalent 175.54 (c) (2) (iii)

MAMMOGRAPHIC INSTALLATIONS

2	1	5414	Lower edge of shield is more than 3 in (7.5cm) from the floor 175.54 (3) (i) (A) (a)
2	1	5415	Shielding does not extend to a height of 6ft (2m) 175.54 (3) (i) (A) (a)
2	1	5416	An air gap exists between the shield and the mammographic unit 175.54 (3) (i) (A) (a)
2	1	5417	Operator is not able to stand completely within the shielded area during the exposure. 175.54 (3) (i) (A) (a)
2	2	5418	Exposure control is not permanently fixed on the mammographic control console 175.54 (3) (i) (A) (b)
2	1	5419	The Operator is not able to communicate with and / or view the patient from operator's protected position during the exposure 175.54 (3) (i) (A) (c)

GENERAL REQUIREMENTS FOR RADIATION EQUIPMENT 175.56

2	2	5601	Radiation equipment not intended to be used was not made inoperable by dismantling or sealing with an Official Department Seal 175.56 (b)
1	1	5602	Radiation equipment unsealed without prior authorization by the Department 175.56 (b)
1	1	5603	Radiation equipment restored to operable condition without prior authorization by the Department 175.56 (b)
2	1	5604	Film processing materials and techniques are not those recommended by the x-ray film manufacturer 175.56 (e)
2	1	5605	Quality control methods not employed to ensure maximum information content of the developed x-ray film 175.56 (e)
2	1	5606	The half- value layer (HVL) of the useful beam is less than the appropriate value shown in Table 1 of this section of the code 175.56 (f) (i)

KUP	Mm Al FILTER
30	0.3
40	0.4
49	0.5
50	1.2
60	1.3
70	1.5
71	2.1
80	2.3
90	2.5
106	2.7
110	3
120	3.2
130	3.5
140	3.8

**DIAGNOSTIC RADIOGRAPHY (other than veterinary) 175.57
(Unit Manufactured after November 28, 1984)**

2	1	5701	The technique factors to be used during an exposure are not indicated on control panel 175.57 (b) (l)
2	1	5702	Indication of technique factors not visible from the operators position 175.57 (b) (l)
1	1	5703	Means not provided to terminate the exposure at a preset time interval 175.57 (b) (2)
2	1	5704	Operator is not able to terminate the exposure at any time during an exposure of greater than one-half second 175.57 (b) (2) (i)
2	1	5705	Exposure can be made when the timer is set to a zero or off position 175.57 (b) (2) (i)
2	1	5706	No indication on control panel when the automatic exposure control mode of operation is selected 175.57 (b) (3) (i)
2	2	5707	A visible signal does not indicate when an exposure has been terminated, when an automatic exposure control provided 175.57 (b) (3) (iv)
2	2	5708	Deviation of technique factors from indicated values exceeds the limits given in the information provided by the manufacturer 175.57 (c)
2	1	5709	For any specific combination of selected technique factors, the estimated coefficient of variation of radiation exposures is greater than 0.05. (exposures are not reproducible) 175.57 (d)
2	1	5710	The average ratios of exposure to the indicated milliamperere – seconds product, obtained at any two consecutive tube current settings differs by more than 0.10 times their sum (exposure are not linear) 175.5 (e) (l)
2	1	5711	The average ratios of exposure to the indicated milliamperere- seconds product obtained at any two consecutive mass selected settings differs by more than 0.10 times their sum (for equipment manufactured after May 3, 1994) (exposures are not linear) 175.57 (e) (2)
2	2	5712	The total misalignment of the edges of the visually defined field with the respective edges pf the x-ray field exceeds 2 percent 175.57 (f) (2) (i) code 175.56 (f) (i)
2	2	5713	The light localizer of the adjustable collimator provides an average illuminance of less than 160 lux (15 foot candles) 175.57 (f) (2) (ii)
3	2	5714	Means not provided to indicate the SID to within 2 percent 175.57 (g) (l)
3	2	5715	Means provided to indicate the SID are not accurate to within 2 percent 175.57 (g) (l)
3	2	5716	The beam – limiting device does not numerically indicate the field size in the plane of the image receptor 175.57 (g) (2)

3	2	5717	X-ray field dimensions do not correspond to those indicated by the beam- limiting device to within 2 percent of the SID 175.57 (g) (3)
2	1	5718	Single purpose equipment not provided with means to limit the field at the plane of the image receptor to dimensions no greater then those of the image receptor 175.57 (h) (l)
3	2	5719	The positive beam limitation system (PBL) does not prevent x-ray production when the beam limiting device is at an SID for which PBL is not designed 175.57 (i) (l) (iii)
3	2	5720	Positive beam limitation system (PBL) is not functioning properly 175.57 (i) (l)
3	2	5721	System does not return automatically to PBL function after a change in image receptor size or SID 175.57 (i) (3)
3	2	5722	No key is available to override the positive beam limiting system (PBL) where the capability for overriding is provided 175.57 (i) (4)
3	2	5723	The key provided for overriding the positive beam limiting system (PBL) can be removed while the positive mode is overridden 175.57 (i) (4)
2	2	5724	The override mode is being used for purposes other than failure of the positive beam limiting system or to perform procedures which cannot be done in the PBL mode 175.57 (i) (4)
3	2	5725	Key provided to override the positive beam limitation systems (PBL) is not properly labeled 175.57 (i) (4)
2	1	5726	Spot – film device is not operating properly 175.57 (j)
2	2	5727	No means provided for mobile or portable x-ray unit to ensure that the minimum source-to- skin distance is at least 30cm (12 in) 175.57 (k)
2	1	5728	No visual indication on x-ray control whenever x-rays are produced 175.57 (l)
2	1	5729	No audible signed is provided to indicate that the exposure has terminated 175.57 (l)
2	1	5730	No means are provided at the control panel and at or near the tube housing assembly, to indicate which tube is being selected, when there are two or more tubes activated by a common exposure switch 175.57 (m)

DENTAL 175.58

2	1	5801	Source to skin is less than 17.8 cm (7 in) for units operating above 50Vp 175.58 (b) (l)
2	1	5802	Source to skin distance is less than 10.2 cm (4 in) for units operating at 50kVp 175.58 (b) (l)

1	1	5803	Beam diameter at the minimum SSD is more than 7 cm (2.75 in) (units manufactured after August 1, 1994) (b) (2) (i)
1	1	5804	Beam diameter at the minimum SSD is more than 7.6 cm (3 in) (units manufactured prior to August 1, 1994) 175.58 (b) (2) (i)
2	1	5805	The operator does not position the end of the position indicating device (PID) within 1 cm (0.4 in) of the skin of the patient 175.58 (b) (2) (ii)
1	1	5806	Device not provided to terminate the exposure after a present time interval or exposure 175.58 (b) (3) (i) (A)
1	1	5807	Exposure switch is not of the dead- man type 175.58 (b) (3) (i) (A)
1	1	5808	Operator cannot terminate the exposure at any time 175.58 (b) (3) (i) (B)
2	1	5809	Termination of the exposure does not cause automatic resetting of the timer to its initial setting or to “zero” 175.58 (b) (3) (i) (D)
2	1	5810	A timer setting beyond the necessary exposure time is being used 175.58 (b) (3) (i) (D)
2	1	5811	The x-ray control does not provide visual indication whenever x-rays are produced 175.58 (b) (3) (ii)
2	2	5812	The visual x-ray production indicator located on the x-ray control is not observable at or from the operator’s protected position 175.58 (b) (3) (ii)
2	1	5813	A signal audible to the operator which indicates that the exposure has terminated is not provided 175.58 (b) (3) (ii)
2	2	5814	Timer operates erratically in that preset exposure times are not reproducible 175.58 (b) (4)
2	2	5815	Deviation of technique factors from the indicated values exceeds the limits specified by the manufacturer 175.58 (b) (5)
2	2	5816	Deviation of technique factors exceeds 10% of the indicated values 175.58 (b) (5)
1	1	5817	Equipment used to make diagnostic dental radiographs of humans has a nominal fixed 50 kVp
1	1	5818	Patient film holding device is being used in a manner which produces a beam diameter at the patients skin of more than 7 cm (2.75 in) (units manufactured after August 1, 1974) 175.58 (b) (7) (i)
1	1	5819	Patient film holding device is being used in a manner which produces a beam diameter at the patients skin of more than 7.6 cm (3 in) (units manufactured prior to August 1, 1974) 175.58 (b) (7) (i)
2	1	5820	The tube housing and / or the PID were hand-held during an exposure 175.58 (b) (7) (ii)

2	1	5821	Time – temperature techniques or automatic processing were not employed to develop pre- operative diagnostic dental x-ray films 175.58 (b) (7) (iv)
2	2	5822	Processing techniques were not consistent with those recommended by the x-ray film manufacturer 175.58 (b) (7) (iv)
2	1	5823	Sight Developing of dental radiographs was being performed 175.58 (b) (7) (iv)
1	1	5824	Excessive patient exposure due to the utilization of improper exposure technique factors 175.58 (b) (7) (v)
1	1	5825	Excessive patient exposure due to the utilization of improper processing conditions 175.58 (b) (7) (v)
1	1	5826	The tube does not remain stationary when placed in the clinical exposure position 175.58 (b) (7) (vii)
2	1	5827	X-ray unit is not equipped with electronic means (timer) for exposure control. (In effect after January 1, 1997) 175.58 (b) (7) (viii)
2	1	5828	Persons not required for the dental x-ray procedure are in the dental x-ray room during the exposure. 175.58
1	1	5829	For extra oral dental radiography, (the x-ray film did not show substantial evidence of cut-off (beam delineation) 175.58 (b) (8)

CERTIFIED DENTAL UNITS ONLY

2	1	5830	The estimated coefficient of variation of radiation exposures is greater than 0.05 for any specific combination of selected technique factors 175.58 (b) (9)
2	1	5831	The average ratios of exposure to the indicated mAs product obtained at any two consecutive tube current settings differs by more than 0.10 times their sum. 175.58
2	1	5832	The minimum half-value layer is less than 1.5 mm aluminum equivalent (units manufactured after December 1, 1980) 175.58 (b) (9) (iv)
2	1	5833	The half-value layer (HVL) of the useful beam is less than the appropriate value shown in Table 1 of this section of the code 175.56 (f) (l)

POADIATRIC RADIOGRAPHY 175.59

1	1	5901	Radiographic beam size is not limited to the area of clinical interest 175.59 (a) (l)
1	1	5902	No substantial evidence of cut- off (beam delineation) on x-ray films 175.59 (a) (2)

1	1	5903	No suitable means are provided to terminate the exposure of podiatric x-ray at a preset time interval or exposure 175.59 (a) (3)
1	1	5904	No adequate dead-man exposure switch provided 175.59 (a) (3)
1	1	5905	Exposure switch cord is not of sufficient length to allow the operator to stand at least 2 meters (6 feet) from the patient and tube and well away from the useful beam during the exposure 175.59 (a) (4)
1	1	5907	No Proper shielded area is provided for operator protection 175.59 (a) (4)
2	1	5908	Person(s) other than required for the x-ray procedure is (are) in the x-ray room during exposure 175.59 (b) (2)
2	1	5909	The half-value layer (HVL) of the useful beam is less than the appropriate value shown in Table 1 of this section of the 175.56 (f) (1)
2	1	5910	Radiation installation did not maintain radiation exposures as far below the limits set forth in this Article as practicable – tube head does not remain stationary when placed in the clinical exposure position 175.51 (n) (l)

FIXED RADIOGRAPHY (EXCLUDING DENTAL, VETERINARY AND PODIATRIC RADIOGRAPHY) 175.60

1	1	6001	Radiographic beam size is not limited to the area of clinical interest 175.60 (a) (l)
1	1	6002	No substantial evidence of cut-off (beam delineation) on x-ray films 175.60 (a) (2)
2	1	6003	General purpose equipment not equipped with an adjustable collimator that visually defines the entire field 175.60 (a) (2)
2	2	6004	The misalignment of the x-ray field with the visual field in either the length or width dimension exceeds 2 percent of the SID (source to image distance) 175.60 (a) (2)
2	2	6005	Legend of the adjustable collimator is not properly calibrated in terms of the projected useful beam size at specified source-film distances 175.60 (a) (2)
2	2	6006	No means provided on the adjustable collimator to numerically indicate the useful beam size at specified source-film distances 175.60 (a) (2)
2	2	6007	No adequate means provided to adjust the projected useful beam size to its numerical indications as noted on the legend of the collimator 175.60 (a) (2)
2	2	6008	The technique factors to be used during an exposure not indicated before the exposure begins 175.60 (a) (3)

2	2	6009	Indication of technique factors is not visible from the operators protected position 175.60 (a) (3)
1	1	6010	No suitable means are provided to terminated the exposure after a preset time interval or exposure 175.60 (a) (4)
2	1	6011	Exposure can be made without using a preset time (timer operates at zero setting) 175.60 (a) (4)
1	1	6012	No adequate dead-man exposure switch 175.60 (a) (5)
2	1	6013	Exposure switch is operable outside the shielded area 175.60 (a) (5)
1	1	6014	No proper shielded area is provided for operator protection 175.60 (a) (5)
1	1	6015	Persons not authorized by the New York City Health Code are holding patient(s) and / or films during x-ray exposures 175.60 (b) (1)
1	1	6016	Individuals occupationally exposed to radiation are being used to hold patients and / or films during x-ray exposures 175.60 (b) (1)
2	2	6017	Mechanical supporting or restraining devices are not available 175.60 (b) (1)
1	1	6018	Appropriate shielding devices are not provided and / or worn by holder(s) of patient(s) 175.60 (b) (1)
1	1	6019	Exposure of an individual used for holding patients is not being monitored 175.60 (b) (1)
2	1	6020	Persons other than required for the radiographic procedure are in the x-ray room during the exposure 175.60 (b) (2)
2	1	6021	Persons, other than the patient, in the room during the radiographic procedure not equipped with appropriate shielding devices 175.60 (b) (2)
1	1	6022	Gonadal shielding of adequate lead equivalency is not being used on patients who have not passed the reproductive age 175.60 (b) (3)

**PORTABLE OR MOBILE (EXCLUDING DENTAL, PODIATRIC,
VETERINARY)
175.61**

1	1	6101	No collimating devices being used to restrict the useful beam to the area of clinical interest 175.61 (a) (1)
1	1	6102	Radiographic beam size is not limited to the area of clinical interest 175.61 (a) (1)
1	1	6103	No substantial evidence of cut-off (beam delineation) on x-ray films 175.61 (a) (2)
2	2	6104	No adequate means provided, for general purpose mobile equipment, to visually define the entire x-ray field 175.61 (a) (2)

3	2	6105	Misalignment of the visually defined field with the x-ray field exceeds 2 (two) percent of the SID (source image receptor distance) 175.61 (a) (2)
3	2	6106	Legend of the adjustable collimator is not properly calibrated in terms of the projected useful beam size at specified source-film distances 175.61 (a) (2)
2	1	6107	Exposure can be made without using a preset time (timer operates at zero setting) 175.61 (a) (3)
2	2	6108	No equipment provided to ensure that the minimum source-to-skin distance is at least 31 cm (12 in) 175.61 (a) (4)
1	1	6109	Persons not authorized or sanctioned by the New York City Health Code are holding patients and / or films during x-ray exposures 175.61
2	1	6110	Mechanical and / or supporting devices are not provided and / or worn by holders of patients and / or films 175.61 (b) (l)
1	1	6111	Appropriate shielding devices are not provided and / or worn by holders of patients and / or films 175.61 (b) (1)
2	1	6112	Parts of the holder's body are in the path of the useful beam 175.61 (b) (l)
1	1	6113	Personnel monitoring devices are not provided and / or worn by holders of patients 175.61 (b) (l)
1	1	6114	Pregnant women used to hold patients during x-ray procedures 175.61 (b) (l)
1	1	6115	Individuals under 18 years of age used to hold patients during x-ray procedure 175.61 (b) (l)
1	1	6116	Gonadal shielding of adequate lead equivalency not being used on patients 175.61 (b) (2)
2	1	6117	Appropriate personnel monitoring is not supplied and / or is not used by persons operating mobile or portable x-ray equipment 175.61 (b) (3)
2	1	6118	Operator not provided with the means to initiate an x-ray exposure at least 2m (6ft) from the patient or behind a protective barrier 175.54 (c) (2) (i)
2	1	6119	Mobile and portable x-ray systems used as fixed radiographic installations do not meet the operator protection standards of 175.54 (l) (i) (A) or (B)
2	1	6120	Operator of a mobile or portable x-ray unit not provided with personnel monitoring 175.54 (c) (2) (ii)
2	1	6121	Operator of a mobile or portable x-ray unit not provided with a protective apron of at least 0.25 mm lead equivalent 175.54 (c) (2) (iii)

1	1	6122	Tubehead does not remain stationary when placed in an exposure position 175.03 (n) (l)
----------	----------	-------------	---

FLUOROSCOPIC 175.62

1	1	6201	The entire cross section of the useful beam is not intercepted by the primary protective barrier 175.62 (a) (l)
1	1	6202	The fluoroscopic tube produces x-rays when the barrier is not in position to intercept the entire useful beam 175.62 (a) (l)
1	1	6203	Exposure rate through barrier with the attenuation block in the useful beam exceeds 2mR/hour at 10 cm (4 in) for each R/min of entrance exposure rate 175.62 (a) (l)
2	2	6204	The misalignment of the edges of the fluoroscopic x-ray field, along any one dimension, exceeds 3% of the SID 175.62 (b) (l)
2	2	6205	The total misalignment along any two dimensions of the fluoroscopic x-ray field exceeds 4% of the SID 175.62 (b) (l)
2	2	6206	The minimum field size, of the fluoroscopic x-ray field, at the greatest SID was not equal to, or less than, 5x5 cm (2x2 in) 175.62 (b) (l)
2	1	6207	During serial fluoroscopy, the operator is unable to terminate the x-ray exposures at any time 175.62 (c)
1	1	6208	The average exposure rate, as measure with a patient phantom type 1100 aluminum, exceeded 5R/min 175.62 (d)
1	1	6209	The maximum exposure rate in air, measured at a point where the useful beam enters the patient, exceeded 10R/min 175.62 (d)
1	1	6210	Maximum exposure rate, for equipment with high option control, exceeded 5R/min when option was not in use 175.62 (d) (l)
1	1	6211	When high option control is in use, an audible signal is not heard 175.62 (d) (l)
1	1	6212	Maximum exposure rate for equipment without automatic exposure rate, and with high option control, exceeded 5R/min when option was not in use 175.62 (d) (2)
1	1	6213	When high option control is in use, for equipment without automatic exposure, an audible signal is not heard 175.62 (d) (2)
2	2	6214	During fluoroscopy tube potential (kV) not continuously indicated 175.62 (e)
2	2	6215	During the fluoroscopy tube current (mA) was not continuously indicated 175.62 (e)
2	2	6216	Tube potential exceeded maximum deviation as stated by the manufacturer 175.62 (e)

2	2	6217	Tube current exceeded maximum deviation as stated by the manufacturer 175.62 (e)
2	1	6218	No means provided to limit the source-skin distance on a stationary fluoroscope to not less than 38 cm (15 in) 175.62 (f)
2	1	6219	No means provided to limit the distance on a mobile fluoroscope to not less than 30 cm (12 in) 175.62 (f)
1	1	6220	The source-skin distance was less than 20 cm (8 in) for a fluoroscope intended for surgical use 175.62 (f)
1	1	6221	Fluoroscopic equipment did not have a preset timer 175.62 (g)
2	1	6222	The maximum cumulative time of the fluoroscopic timing device exceeded 5 minutes without resetting 175.62 (g)
2	2	6223	No signal, audible to the fluoroscopist, to indicate completion of any preset cumulative time, was present 175.62 (g)
2	2	6224	After preset time expired, fluoroscopic signal did not continue to sound, while x-rays were produced, before timer was reset 175.62 (g)
2	1	6225	High contrast resolution of the fluoroscopic system was not capable of resolving a minimum mesh number of 24 , for the center of the beam and 20 for the edges, using a test tool described in this section 175.62 (h)
2	1	6226	Low contrast performance to the fluoroscopic system was not capable of resolving a minimum hole size of 3 mm , using a test tool described in this section 175.62 (h)
2	1	6227	Fluoroscopic patient rates were not determined annually 175.62 (i) (l)
3	1	6228	Fluoroscopic patient rates were not determined when there was a major component replaced or serviced 175.62 (i) (l)
2	1	6229	Measurements of stray radiation, to operators and / or observers, were not performed when equipment was first placed in operation 175.62 (i) (l)
2	1	6230	Fluoroscopic patient rates were not determined when equipment was first placed in operation 175.62 (i) (l)
1	1	6231	Protective garments of at least 0.25 mm lead equivalent were not worn by fluoroscopist during every fluoroscopic examination 175.62 (i) (2)
1	1	6232	Protective gloves and aprons of at least 0.25 mm lead equivalent were not worn by persons within the fluoroscopy room 175.62 (i) (3)
1	1	6233	Persons not needed in the fluoroscopy room were present during exposures 175.62 (i) (4)